



68U399 Electrode Zero Point Calibrator



This electrode position calibrator is used to set the Zero Position for electrode in stereotaxic surgery. In the vertical direction, the reference point is 30mm higher

than that of ear bars. It was designed for 68U000 stereotaxes.