



CARFIL QUALITY
YOUR SUPPLIER FOR ALL LABORATORY NEEDS

68U397 Level Calibrator - Rat



The 68U397 calibrator refers to the international brain atlas of rats and takes craniofacial bregma and lambda as the reference points. The distance between B(0) and L(0) is about 7,67mm. When the animal craniofacial is in the horizontal state, these two points should be on the same level. When fixed to the X-Z manipulator of stereotaxic instrument and descends, the tip of B(0) and L(0) should touch to craniofacial at the same time. Otherwise the adaptor should be adjusted to meet the requirements.



PAVAN SERVICE BVBA

Beyntellus 3 - B-2360 Oud-Turnhout - Tel. +32 (0)14 45 13 10 - admin@carfil.be - BTW BE 0420.681.872

www.carfil.be