



CARFIL QUALITY
YOUR SUPPLIER FOR ALL LABORATORY NEEDS

Pulse Oximeter VET Handheld Pulse Oximeter (sensors)



Sensors

Purelight Vet sensors produce the pure light spectrum which eliminates variations in readings from animal-to-animal and sensor-to-sensor.

2000SL -Clip-on sensor for tongue applications, paw pads, and well-vascularized areas.

2000T -For placement on the underside base of the tail or other well-vascularized surfaces.

2000SA -Wrap sensor for placement on the toe (large animal) or base of the tail or foot (small animal).

Sensor Placement

2000SL Lingual Clip Sensor

Is convenient for spotchecks and for monitoring during recovery while the animal is immobile. However the sensor is held in place by light spring pressure that could allow it to become dislodged with movement. Further, over time even light spring compression may interfere with blood flow resulting in signal loss and need to reposition the sensor. The 2000SL is most easily applied to the rear foot, but the front foot and tail can also be used. For mice, the sensor can be best placed high into the groin.



2000T Transflectance Sensor

Is also suitable for continuous monitoring, the 2000T transflectance sensor is the smallest probe and can be applied to the rear or front feet, or to the ventral surface of the tail of the rat. Secured with adhesive tape.



Alternatively the sensor can be attached to the surface of an operating or imaging table and the foot simply taped into position over it.



PAVAN SERVICE BVBA

Beyntellus 3 - B-2360 Oud-Turnhout - Tel. +32 (0)14 45 13 10 - admin@carfil.be - BTW BE 0420.681.872

www.carfil.be



CARFIL QUALITY
YOUR SUPPLIER FOR ALL LABORATORY NEEDS

2000SA Small Animal Wrap Sensor

Can be easily secured and less prone to accidental displacement, making it an ideal option for continuous monitoring during long surgical or other procedures. Adhesive tape, cohesive bandage or a Posey Wrap can be used to assure that the sensor diodes are aligned directly opposite each other and that excessive pressure is avoided. The 2000SA should not be applied to hair-covered or highly pigmented areas.



Ordering number

2000SL	Sensor Lingual clip Sensor
2000T	Transflectance sensor
2000SA	Small Animal Wrap



PAVAN SERVICE BVBA

Beyntellus 3 - B-2360 Oud-Turnhout - Tel. +32 (0)14 45 13 10 - admin@carfil.be - BTW BE 0420.681.872

www.carfil.be